


---

# MicroBooNE status

AEM – 9<sup>th</sup> Nov 2015


# DAQ uptime

- DAQ uptime average over 97% for the last week




# POT


- Total POT written to tape this week: 0.995e19
- POT efficiency of 97.8%



POT (Week of 11/2/2015)




POT



# Trigger efficiency

- Low (85%) NuMI trigger efficiency observed
  - Compared to >99% for BNB
- Discovered problem with logic at LArTF
- Efficiency now at 98% since Friday
- Expect this to increase to 99.5% by removing the dependency on \$AE signals



# BNB downtime

---

- BNB downtime today, and possibly tomorrow
- This downtime is being used for studies which will be used to commission our PMT trigger
- Also migrating online software repositories
- Will be ready for beam when it comes back